LADY READING HOSPITAL – MEDICAL TEACHING INSTITUTION PESHAWAR

AFTER GREVIANCE REDRESSAL COMMITTEE BID EVALUATION REPORT (BER) PATHOLOGY AND BLOOD BANK PROPRIETARY FY 2024-25

Bids evaluation technical score is hereby widely shared through the official website of LRH-MTI in public interest with the following explanations for further necessary action by the concerned bidders:

- 1. The list attached is a draft tentative Bid Evaluation Report (BER) evaluated by technical evaluation (T&E) committee to serve the only purpose of voluntary disclosure of information to facilitate ensuring all the more transparency and equity in the bidding process by providing equal opportunities to all concerned bidders by enabling them to have access to information related to their bids in the fairness of justice and competition;
- 2. This list will be kept on this official website till 9:00 a.m. 10/01/2025; and
- 3. Although every possible effort has been made to prepare this draft tentative BER, the same is being shared to further eliminate the possibility of any error, omission in terms of any details related to the technical evaluation of items quoted by the concerned bidders.
- 4. Any feedback related to errors and omissions must reach the office of Manager Material Management, LRH-MTI Peshawar not later than

9:00 a.m. 10/01/2025; in written form duly signed and attested by the responsible relevant person of the bidding entity.

4. Any comments, feedback etc. received after the above mentioned cutoff time and date (Para-4 above) due to any reason, whatsoever, will not be considered and the so finalized bid evaluation report (BER) will be placed before the purchase Committee, LRH-MTI Peshawar, for its approval in the best public interest.

The summary of bidders and their status is given as;

				machine in				Importer.	ed in any gation		ne Certifica ificate is manda	cer Cer	Firm rtification	Relevent Experience	Satisfactory Performance Certificate of Firm		ualified mical Sta		Fina	anical	Capality		/are ouse		
s	. No Supplier	Model	Country of Origin	Machine Specification according to already placed Hospital	NTN Registration Certificate	Sales Tax Registration Certificate	KNTN registration Certificate	ate from th	Undertaking that Firm/Company is not blacklisted in ang government/Seni Government and have no litigation pending in any court of law in Pakistan	US Food and Drug Administration	- J	Internation of the Industrial Standard	9001 F	I. 01 - 03 Years Market Experience (Justified by Purchase/Supply/Work Orders) II. 04 - 06 Years	Satisfactory Performance certificate from different Institution, Maximum 4 Certificate 02 Marks for 1 Certificate.	Diploma Engineer	Graduate Engineers Bio-Medical/Electronics	Foreign Trained Engineer Maintenance/Services Machine uptime will be within 12 Hours. Penalty of Rs. 2,000. per Hour will be charged in case of down time exceed from 12 hours in case if backup is in working order. In case if both are out of order penalty is after 02 hours. Confirmation through Letter Head.	Turnover in Million (I-T-I and IT-II Form) i. 01-50 Million	ii. 50-100 Million	Bank Statement of Last 02 Years (2.5 Marks for each year) Audit Report of last 2 years (2.5 Marks for each	year) Ware house at Peshawar	House at National Le	Technical Marks	Remarks
					Mandatory	Mandtory	2 Ma	andatory !	Mandatory	2		2	2	2 2	8	2	3	3 5	4	6	5 5	10	5	70	

Kits and consumables with the placement of fully automated

Arterial Blood Gas System (Fully advanced latest system) for the measurement of arterial blood gasses. (Roche, Abbot, Siemens or Equivalent with latest version and higher specifications than the existing instrument/equipment/analyzer)

Sure Bio Diagnostics & Pharmaceuticals	PT1000 Blood Gas & Electrolyte Analyzer	China										QOUT	ED CHINA	ORIG	GION											Non-Responsive
2 Meditec Instruments Co.	Easystat	Medica USA	Not Approved	~	~	0	· •	,	2	0	0	0	2	2	8	2	3	0	5	0	0	5	5	10 5	49	Non-Responsive Due to Multiple Parameters Missing
3 Martin Dow Specialities Pvt Ltd.	Prime CCS	USA	Not Approved	1	1	0	· •	,	2	2	0	2	2	2	8	2	3	3	5	4	6	5	5	10 5	66	Non-Responsive Due to Multiple Parameters Missing
4 Burhani Enterprises	Techno Medica Model: GASTAT-735)	Japan	Approved	*	1	2	· •	,	0	2	2	2	2	2	8	2	3	3	5	4	6	5	5	10 5	68	Non-Responsive Due to Urea Parameters Missing

5	Hoora Pharma	Rapid Lab 1200	Siemens Healthineers USA	Not Approved	1	1	2	✓	~	2	2	0	2	2	2	8	2	3	3	5	4	6	5	5	10 5	5 68	Non-Responsive Due to Urea Parameters Missing
6	Roche Pakistan Ltd	Cobas b 221	Hitachi Japan	Approved	1	✓	0	✓	~	2	2	0	0	2	2	8	2	3	3	5	4	6	5	5	0 5	5 54	Responsive
	Kits and Consumables with The Placement of Fully Automated Chemistry Analyzer (Fully advanced latest version) (Roche, Abbot, Siemens or Equivalent with latest version and higher specifications than the existing instrument/equipment/analyzer) Hoora Pharma Attilica CH 930 Siemens Healthineers Approved																										
1	Hoora Pharma	Attilica CH 930	Siemens Healthineers USA	Approved	4	1	2	✓	~	2	2	0	2	2	2	8	2	3	3	5	4	6	5	5	10 5	5 68	Responsive
2	Roche Pakistan Ltd	Cobas C 303	Hitachi Japan	Approved	1	1	0	1	1	2	2	0	0	2	2	8	2	3	3	5	4	6	5	5	0 5	5 54	Responsive
	Kits & Consumables with Placement of Fully Automated Immunodiagnostic System (Special Chemistry) Fully advanced latest version (Roche, Abbot, Siemens or Equivalent with latest version and higher specifications than the existing instrument/equipment/analyzer) Hoora Pharma Attilica IM 1600 Siemens Healthineers USA Approved V V 2 V V 2 0 2 0 2 2 8 2 3 3 5 4 6 5 5 10 5 68 Responsive																										
1	Hoora Pharma	Attilica IM 1600		Approved	1	1	2	1	1	2	2	0	2	2	2	8	2	3	3	5	4	6	5	5	10 5	68	Responsive
2	Roche Pakistan Ltd	Cobas E 402	Hitachi Japan	Approved	1	1	0	✓	~	2	2	0	0	2	2	8	2	3	3	5	4	6	5	5	0 5	5 54	Responsive
	Kits and Consumables with The Placement of Fully Automated Hematology Analyzer (Fully advanced latest version-7 parts or above) (Roche, Abbot, Siemens, Backman, Nihon Kohden or Equivalent with latest version and higher specifications than the existing instrument/equipment/analyzer) Estimated workload/year S. Ejazuddin & Co Sysmex XN-1000 (XN-10) USA Approved																										
1	S. Ejazuddin & Co	Sysmex XN-1000 (XN-10	USA	Approved	✓	1	2	✓	1	2	2	2	2	2	2	8	2	3	3	5	4	6	5	5	10 5	70	Responsive
2	NexTek Health Care	DxH 900 Beckman Coulte	er Ireland	Approved	1	1	2	✓	✓	2	2	0	2	2	2	8	2	3	3	5	4	6	5	5	10 5	5 68	Responsive
3	Burhani Enterprises	Abbott Alinity-H	Abbott USA	Approved	✓	1	2	✓	✓	2	2	0	2	2	2	8	2	3	3	5	4	6 :	5	5	10 5	68	Responsive
	Kits and Consumables with The Placement of Fully Automated HPLC/Capillary (Fully advanced latest version or above) Test Range 20 or more With All specifications and higher specifications than the existing instrument/equipment with controls, calibrators, buffers UPS, Computer system etc.																										
1	Sure Bio Diagnostics & Pharmaceuticals	H8 HbA1c HPLC Analyze	er China										QOUT	ED CHINA	ORIGIO	N											Non-Responsive
									Kits and y Automate th All specifi	ESR (F	ully ad		st version	or above)													
1	Burhani Enterprises	ESR Analyzer (Model: Cube-30 Touch)	Diesse Diagnostic Evolution/Italy	Approved	4	1	2	~	~	2	2	0	2	2	2	8	2	3	3	5	4	6 :	5	5	10 5	5 68	Responsive
					Fully A	utomate	d FLO	WCYT	OMETRY (Fully ad	lvanced	with The P l latest vers ens, Guava	sion or ab	ove with h	igher nui	mber of cha	nnels)										
1												R PARTICIP															
						Fully	Autor	nated G	6PD Quanti	itative P	OCT d	with The P evice (Fully ns, Biosens	y advance	d latest ve	ersion or	above)											
1										NO V	ENDOF	R PARTICIP	PATED														

Kits and Consumables with The Placement of Fully Automated Machine Blood Bank (Roche, Abbott, Siemens Or Equivalent) 1 Martin Dow Specialities Pvt Ltd. Vitros 3600 Johson & USA Approved																												
1	Martin Dow Specialities Pvt Ltd.		USA	Approved	~	~	0	1	~	2	2	0	2	2	2	8	2	3	3	5	4	6	5	5	10	5	66	Responsive
2	Burhani Enterprises	Architect 2000	Abbott Core Laboratory/USA	Not Approved	1	1	2	1	1	2	2	0	2	2	2	8	2	3	3	5	4	6	5	5	10	5	68	Non Responsive due to quoted old Model.
3	Hoora Pharma	Attilica IM 1300	Siemens Healthineers USA	Approved	~	1	2	1	1	2	2	0	2	2	2	8	2	3	3	5	4	6	5	5	10	5	68	Responsive
4	Roche Pakistan Ltd	Cobas E 402	Hitachi Japan	Approved	~	1	0	1	*	2	2	0	0	2	2	8	2	3	3	5	4	6	5	5	0	5	54	Responsive
	Kits and Consumables with The Placement of Fully Automated Machine Blood Bank (Roche, Abbott, Siemens Or Equivalent) Card System for Pre-Transfusion Blood Cross-Match Sure Bio Diagnostics & Polyspecific (Ortho Clinical UK Approved																											
Ge	Card System for Pre-Ti		Cross-Match		I									1						1				_	Т			
1	Sure Bio Diagnostics & Pharmaceuticals	Polyspecific (Ortho Clinica Diagnostic)	al UK	Approved	~	~	2	~	✓	0	2	0	2	2	2	8	2	3	3	0	4	6	5	5	10	5	61	Responsive
Blo	od Group by Gel Cards																											
1	Sure Bio Diagnostics & Pharmaceuticals	ABD Confirmation 1 Card = 2 Blood Grouping (Orth Clinical Diagnostic)		Approved	1	1	2	✓	~	0	2	0	2	2	2	8	2	3	3	0	4	6	5	5	10	5	61	Responsive
Th	ree Cells Panels for Antiboo	dies detection (for r	rare Blood Groun																									
1	Sure Bio Diagnostics & Pharmaceuticals	Cell Screen 1-2-3 PAP (3x5ml) (Redcell Biotechnology)	Tukey	Not Approved	1	1	2	~	1	0	2	0	2	2	2	8	2	3	3	0	4	6	5	5	10	5	61	Non-Responsive
	Dhaymacayticala																											
1	Sure Bio Diagnostics & Pharmaceuticals	Cell ID 11 (11x5ml) (Redcell Biotechnology)	Turkay	Not Approved	~	✓	2	✓	✓	0	2	0	2	2	2	8	2	3	3	0	4	6	5	5	10	5	61	Non-Responsive
Ge 1	Sure Bio Diagnostics & Pharmaceuticals	BioVue Anti-Human Polyspecific (Ortho Clinica	al UK	Approved	✓	~	2	~	✓	0	2	0	2	2	2	8	2	3	3	0	4	6	5	5	10	5	61	Responsive
		Diagnostic)			1	ļ	1				<u> </u>												1	1		_		
Ge	card for Rh antibodies D,	Kn Pnehotype (48	n phenotyping an	d weak D/D-varia	nt																			_	Т			
1	Sure Bio Diagnostics & Pharmaceuticals	Cards/Box) Rh Group System Each Parameter 2 Test (Redcell	2 Turkey	Not Approved	1	~	2	4	1	0	2	0	2	2	2	8	2	3	3	0	4	6	5	5	10	5	61	Non-Responsive
An	ti-A1 Lectin reagent and A	nti-H lectin reagent	t																									
1	Sure Bio Diagnostics & Pharmaceuticals	A1 Gel (48 Cards/ Box) (Redcell Biotechnology)		Not Approved	1	~	2	✓	✓	0	2	0	2	2	2	8	2	3	3	0	4	6	5	5	10	5	61	Non-Responsive
				(Roche A	Abbot Si	emenc	or Fan		and Consu Coagulatio	n Analyz	zer (Fu	lly advan	ced latest v	rersion)		rument/equip	omen t	/apalv	vzer)									
1	Martin Dow Specialities Pvt Ltd.	Satellite Max	USA	Approved	√	✓	0	√ valent v	✓	2	2	0	2	2	2	8	2	3		5	4	6	5	5	10	5	66	Responsive
2	Burhani Enterprises	CA-1600	Sysmex Corporation/Japan	Approved	~	1	2	*	✓	0	2	2	2	2	2	8	2	3	3	5	4	6	5	5	10	5	68	Responsive

				K	its and (Consum	ables v	with The					od Culture s Equivalent)	ystem w	vith UPS	(240 bottle	es capa	acity)											
1	IMCO Technologist	Bactec FX (Capacity 200- 400 Bottles) Becton Dickinson USA	USA	Not Approved	~	~	0	~	1	2	2	0	0	2	2	8		2	3	0	5	4	6	5	:	5	10	5 61	Non-Responsive due to Specification (Machine Quoted of USA Origin and Consumable quoted of China Origin
2	Global Marketing Services	Bact Alert 3D 240 BioMerieux	France	Approved	1	1	2	1	1	0	2	0	2	2	2	8		2	3	3	0	4	6		: :	5	10	61	Responsive
3	Sure Bio Diagnostics & Pharmaceuticals	Automated Blood Culture System (BC60)	China									•	QOUT	ED CHI	NA ORIG	ION			•							•	•		Non-Responsive
4	Burhani Enterprises	BC 32 Render Bioteck	China																									Non-Responsive	

Note: The Blood Culture System will be Re Tender to ensure healthy Competition.

Assistant Manager Material Management Blood Bank Incharge HOD Patholoty